Medikamentenvernebler Set für Kinder, inkl. Maske und O_2-Schlauch

REF IC-VMA01 | Menge: 1 Set | MD | CE 1639

INHALT:

Medikamentenvernebler Set für Kinder, inkl. Maske und 210 cm O_2-Schlauch

INDIKATION

Zur inhalativen Aufnahme von Medikamenten als Aerosol / Behandlung verschiedener Erkrankungen der Atemwege.

ZIELGRUPPE

Pflegepersonal, medizinische Fachkräfte

ANWENDUNGSBESCHREIBUNG

1. Schrauben Sie den Deckel des Medikamentenverneblers ab und füllen das Medikament in den Vernebler-Topf.
2. Schrauben Sie den Deckel wieder auf den Vernebler-Topf.
3. Stecken Sie die Maske in den Medikamentenvernebler-Deckel und verbinden den Vernebler-Topf mittels O2-Schlauch mit der Flowquelle.
4. Stellen Sie den Flow nach Anweisung des Arztes ein.
5. Setzen Sie dem Patienten die Maske auf und fixieren sie mit dem Kopfhalteband.

WARNHINWEISE

- Einmalprodukt, nicht wiederverwenden oder aufbereiten
- Nicht verwenden, wenn die Verpackung bereits geöffnet oder beschädigt ist
- Kühl und trocken lagern
- Achten Sie darauf, dass alle Konnektoren fest sitzen und der Verbindungsschlauch durchgängig ist

HINWEIS ZU MELDEPFLICHTEN

Bitte informieren Sie den Hersteller, den Vertreiber und den Importeur in allen Fällen von schwerwiegenden Ereignissen in Verbindung mit dem Produkt.

6 945099 302465

LOT 210508
2021-07
2026-06

Hersteller | Manufacturer
Ningbo Shengyurui Medical Appliances CO., LTD
No 138, Binhaisi Road, Hangzhou Bay New Zone,
315336 Ningbo, Zhejiang Province P.R. China

EC REP EU Repräsentant | EC Representative
Shanghai International Holding Corp. GmbH (Europe)
Eiffestrasse 80, 20537 Hamburg
info@shholdingeu.com

Importeur | Importer
Novo Klinik-Service GmbH
Technologiepark West - Haus 15
Zum Frenser Feld 1 | 50127 Bergheim
Germany
Tel.+49 2271 99404-0 | Fax +49 2271 99404-100
info@its-novo.de | its-novo.de

Revison: 01-20210112-IFU-ICVMA01

Nebulizer set for children, incl. mask and oxygen sure flow tubing

REF IC-VMA01 | Quantity: 1 Set | MD | CE 1639

CONTENT:

Nebulizer set for children, incl. mask and 210 cm oxygen sure flow tubing

INDICATIONS

Mainly used for the inhalation of medication as aerosol / in the treatment of various diseases of the respiratory system.

INTENDED USERS

Nursing staff, medical professionals

APPLICATION DESCRIPTION

1. Place prescribed medication in nebulizer jar by unscrewing lid.
2. Reattach lid.
3. Insert the mask into the nebulizer and connect the nebulizer to the flow source using an O2 tube.
4. Set liter flow as prescribed by physician.
5. Place the mask on the patient and secure the mask with the head strap.

WARNING NOTICES

- For single use. Discard after use. Do not reprocess.
- Do not use if package is open or damaged
- Store in cool and dry place
- Ensure that all connections are secure and air circulates freely through the tubing

NOTICE ON REPORTING REQUIREMENTS

Please inform the manufacturer, the distributor and the importer in all cases of serious incidents in connection with the product.

MD Medizinprodukt / Medical device
Gebrauchsanweisung beachten / Consult instructions for use
Nicht wiederverwenden / Do not re-use
Von Sonnenlicht fernhalten / Keep away from sunlight
Trocken lagern / Keep dry
Bei beschädigter Verpackung nicht verwenden / Do not use if package is damaged
Achtung / Caution

Hersteller | Manufacturer
Ningbo Shengyurui Medical Appliances CO., LTD
No 138, Binhaisi Road, Hangzhou Bay New Zone,
315336 Ningbo, Zhejiang Province P.R. China

EC REP EU Repräsentant | EC Representative
Shanghai International Holding Corp. GmbH (Europe)
Eiffestrasse 80, 20537 Hamburg
info@shholdingeu.com

Importeur | Importer
Novo Klinik-Service GmbH
Technologiepark West - Haus 15
Zum Frenser Feld 1 | 50127 Bergheim
Germany
Tel.+49 2271 99404-0 | Fax +49 2271 99404-100
info@its-novo.de | its-novo.de

Revison: 01-20210112-IFU-ICVMA01